

## Electronic Weights & Measures Seal

as

appendix to measurement system certificate

Internal state type approval

5.602

97.10

MultiFlow

The seal was produced by:


Signature and identification of official:

--

**It is essential to observe the following instructions when checking the seal status:**

- The seal is not violated by the inspection.
- The seal impression opposite for the Multiflow can be repeated with the following key combination:  
*Switch On*      <F1> (Seal Status)  
                         <F1> (Print)
- The code number in the 'Seal status' area on the copy and on the original must match (double-sized printing).
- The text under the code number must read:  
\*The seal is approved!\*
- If the details do not match, the seal has been broken. Appropriate action must be initiated.

**IMPORTANT:**

**Illegal modification of the W & M data or the W & M seal is a punishable offence!**